



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Accuracy Calibration Services S. de R.L. de C.V.***  
***Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamaulipas 88993***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Dimensional, Weighing Devices, Acoustic, Time & Frequency, Mechanical and Electrical Calibration***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Expiration Date:</i>
March 17, 2011	September 28, 2015	September 30, 2017

<i>Accreditation No.:</i>	<i>Certificate No.:</i>
69876	L15-315

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Current <sup>FO</sup>	4.3 $\mu$ A to 3.299 99 mA	0.13 % of reading + 0.05 $\mu$ A	Fluke 5500A
	3.3 mA to 32.999 9 mA	0.01 % of reading + 0.25 $\mu$ A	
	33 mA to 329.999 mA	0.01 % of reading + 3.3 $\mu$ A	
	330 mA to 2.199 9 A	0.34 % of reading + 44 $\mu$ A	
	2.2 A to 11 A	0.06 % of reading + 330 $\mu$ A	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
10 Hz to 20 Hz	29 $\mu$ A to 0.329 99 mA	0.24 % of reading + 0.15 $\mu$ A	
20 Hz to 45 Hz	29 $\mu$ A to 0.329 99 mA	0.13 % of reading + 0.15 $\mu$ A	
45 Hz to 1 kHz	29 $\mu$ A to 0.329 99 mA	0.13 % of reading + 0.25 $\mu$ A	
1 kHz to 5 kHz	29 $\mu$ A to 0.329 99 mA	0.4 % of reading + 0.15 $\mu$ A	
5 kHz to 10 kHz	29 $\mu$ A to 0.329 99 mA	1.25 % of reading + 0.015 mA	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
10 kHz to 20 kHz	0.33 mA to 3.299 9 mA	0.2 % of reading + 0.3 $\mu$ A	
20 Hz to 45 Hz	0.33 mA to 3.299 9 mA	0.1 % of reading + 0.3 $\mu$ A	
45 Hz to 1 kHz	0.33 mA to 3.299 9 mA	0.1 % of reading + 0.3 $\mu$ A	
1 kHz to 5 kHz	0.33 mA to 3.299 9 mA	0.2 % of reading + 0.3 $\mu$ A	
5 kHz to 10 kHz	0.33 mA to 3.299 9 mA	0.6 % of reading + 0.3 $\mu$ A	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
10 Hz to 20 Hz	3.3 mA to 32.999 mA	0.2 % of reading + 3 $\mu$ A	
20 Hz to 45 Hz	3.3 mA to 32.999 mA	0.1 % of reading + 3 $\mu$ A	
45 Hz to 1 kHz	3.3 mA to 32.999 mA	0.09 % of reading + 3 $\mu$ A	
1 kHz to 5 kHz	3.3 mA to 32.999 mA	0.2 % of reading + 3 $\mu$ A	
5 kHz to 10 kHz	3.3 mA to 32.999 mA	0.6 % of reading + 3 $\mu$ A	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
10 Hz to 20 Hz	33 mA to 329.99 mA	0.2 % of reading + 30 $\mu$ A	
20 Hz to 45 Hz	33 mA to 329.99 mA	0.1 % of reading + 30 $\mu$ A	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			Fluke 5500A
45 Hz to 1 kHz	33 mA to 329.99 mA	0.09 % of reading + 30 $\mu$ A	
1 kHz to 5 kHz	33 mA to 329.99 mA	0.2 % of reading + 30 $\mu$ A	
5 kHz to 10 kHz	33 mA to 329.99 mA	0.6 % of reading + 30 $\mu$ A	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
10 Hz to 45 Hz	0.33 A to 2.199 99 A	0.2 % of reading + 300 $\mu$ A	
45 Hz to 1 KHz	0.33 A to 2.199 99 A	0.1 % of reading + 300 $\mu$ A	
1 KHz to 5 KHz	0.33 A to 2.199 99 A	0.75 % of reading + 300 $\mu$ A	
Equipment to Measure AC Current At the listed frequencies <sup>FO</sup>			
45 Hz to 60 Hz	2.2 A to 11 A	0.06 % of reading + 2 mA	
65 Hz to 500 Hz	2.2 A to 11 A	0.1 % of reading + 2 mA	
500 Hz to 1 KHz	2.2 A to 11 A	0.33 % of reading + 2 mA	
Equipment to Output Resistance <sup>FO</sup>			
	28 m $\Omega$ to 10.99 $\Omega$	9.3 m $\Omega$	
	11 $\Omega$ to 32.999 $\Omega$	19 m $\Omega$	
	33 $\Omega$ to 109.999 $\Omega$	25 m $\Omega$	
	110 $\Omega$ to 329.999 $\Omega$	45 m $\Omega$	
	330 $\Omega$ to 1.099 99 k $\Omega$	0.16 $\Omega$	
	1.1 k $\Omega$ to 3.299 99 k $\Omega$	0.36 $\Omega$	
	3.3 k $\Omega$ to 10.999 9 k $\Omega$	1.6 $\Omega$	
	11 k $\Omega$ to 32.999 9 k $\Omega$	3.6 $\Omega$	
	33 k $\Omega$ to 109.999 k $\Omega$	19 $\Omega$	
	110 k $\Omega$ to 329.999 k $\Omega$	46 $\Omega$	
	330 k $\Omega$ to 1.099 9 M $\Omega$	220 $\Omega$	
	1.1 M $\Omega$ to 3.2999 9 M $\Omega$	550 $\Omega$	
	3.3 M $\Omega$ to 10.999 9 M $\Omega$	7.2 k $\Omega$	
	11 M $\Omega$ to 32.999 9 M $\Omega$	39 k $\Omega$	
	33 M $\Omega$ to 109.999 M $\Omega$	610 k $\Omega$	
	110 M $\Omega$ to 330 M $\Omega$	1.9 M $\Omega$	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Voltage <sup>FO</sup>	0.69 mV to 329.999 9 mV	210 $\mu$ V	Fluke 5500A
	330 mV to 3.299 9 V	170 $\mu$ V	
	3.3 V to 32.999 9 V	1.8 mV	
	33 V to 329.999 9 V	19 mV	
	330 V to 1 020 V	58 mV	
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			
10 Hz to 40 Hz	1 mV to 32.999 mV	135 $\mu$ V	
40 Hz to 10 kHz	1 mV to 32.999 mV	71 $\mu$ V	
10 kHz to 20 kHz	1 mV to 32.999 mV	89 $\mu$ V	
20 kHz to 50 kHz	1 mV to 32.999 mV	0.12 mV	
50 kHz to 100 kHz	1 mV to 32.999 mV	0.34 mV	
100 kHz to 500 kHz	1.35 mV to 32.999 mV	0.45 mV	
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			
10 Hz to 40 Hz	33 mV to 329.99 mV	0.85 mV	
40 Hz to 10 kHz	33 mV to 329.99 mV	0.85 mV	
10 kHz to 20 kHz	33 mV to 329.99 mV	0.35 mV	
20 kHz to 50 kHz	33 mV to 329.99 mV	0.56 mV	
50 kHz to 100 kHz	33 mV to 329.99 mV	2 mV	
100 kHz to 500 kHz	33 mV to 329.99 mV	6.4 mV	
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			
10 Hz to 40 Hz	330 mV to 3.299 9 V	2.6 mV	
40 Hz to 10 kHz	330 mV to 3.299 9 V	2.6 mV	
10 kHz to 20 kHz	330 mV to 3.299 9 V	2.7 mV	
20 kHz to 50 kHz	330 mV to 3.299 9 V	5.0 mV	
50 kHz to 100 kHz	330 mV to 3.299 9 V	9.7 mV	
100 kHz to 500 kHz	330 mV to 3.299 9 V	20 mV	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			Fluke 5500A
10 Hz to 40 Hz	3.3 V to 32.999 9 V	14 mV	
40 Hz to 10 kHz	3.3 V to 32.999 9 V	14 mV	
10 kHz to 20 kHz	3.3 V to 32.999 9 V	29 mV	
20 kHz to 50 kHz	3.3 V to 32.999 9 V	68 mV	
50 kHz to 100 kHz	3.3 V to 32.999 9 V	98 mV	
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			
40 Hz to 1 kHz	33 V to 329.999 V	0.17 V	
1 kHz to 10 kHz	33 V to 329.999 V	0.27 V	
10 kHz to 20 kHz	33 V to 329.999 V	0.32 V	
Equipment to Measure AC Voltage At the listed frequencies <sup>FO</sup>			
40 Hz to 1 kHz	330 V to 1 020 V	0.58 V	
1 kHz to 5 kHz	330 V to 1 020 V	2.2 V	
5 kHz to 10 kHz	330 V to 1 020 V	2.6 V	
Equipment to Measure Capacitance At the listed frequencies <sup>FO</sup>			
0.33 nF to 0.4999 nF	50 Hz to 1 000 Hz	0.012 nF	
0.5 nF to 1.0999 nF	50 Hz to 1 000 Hz	0.015 nF	
1.1 nF to 3.299 nF	50 Hz to 1 000 Hz	0.026 nF	
3.3 nF to 10.999 nF	50 Hz to 1 000 Hz	0.064 nF	
11 nF to 32.999 nF	50 Hz to 1 000 Hz	0.18 nF	
33 nF to 109.99 nF	50 Hz to 1 000 Hz	0.37 nF	
110 nF to 329.99 nF	50 Hz to 1 000 Hz	1.2 nF	
0.33 $\mu$ F to 1.099 9 $\mu$ F	50 Hz to 1 000 Hz	3.8 nF	
1.1 $\mu$ F to 3.299 9 $\mu$ F	50 Hz to 1 000 Hz	15 nF	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure Capacitance At the listed frequencies <sup>FO</sup>			Fluke 5500A
3.3 $\mu$ F to 10.999 $\mu$ F	50 Hz to 400 Hz	49 nF	
11 $\mu$ F to 32.999 $\mu$ F	50 Hz to 400 Hz	270 nF	
33 $\mu$ F to 109.99 $\mu$ F	50 Hz to 200 Hz	650 nF	
110 $\mu$ F to 329.99 $\mu$ F	50 Hz to 100 Hz	2.6 $\mu$ F	
330 $\mu$ F to 1.1 mF	50 Hz to 100 Hz	12 $\mu$ F	
Equipment to Output DC Current <sup>FO</sup>			HP 3458A
	30 $\mu$ A to 100 $\mu$ A	0.002 5 % of reading + 0.8 nA	
	100 $\mu$ A to 1 mA	0.002 5 % of reading + 5 nA	
	1 mA to 10 mA	0.002 5 % of reading + 50 nA	
	10 mA to 100 mA	0.004 % of reading + 500 nA	
	100 mA to 1 A	0.012 % of reading + 10 $\mu$ A	
Equipment to Output Resistance <sup>FO</sup>			HP 3458A
	200 $\mu$ $\Omega$ to 10 $\Omega$	0.001 5 % of reading + 50 $\mu$ $\Omega$	
	10 $\Omega$ to 100 $\Omega$	0.001 2 % of reading + 500 $\mu$ $\Omega$	
	100 $\Omega$ to 1 k $\Omega$	0.001 % of reading + 500 $\mu$ $\Omega$	
	1 k $\Omega$ to 10 k $\Omega$	0.01 % of reading + 5 m $\Omega$	
	10 k $\Omega$ to 100 k $\Omega$	0.01 % of reading + 50 m $\Omega$	
	100 k $\Omega$ to 1 M $\Omega$	0.15 % of reading + 2 $\Omega$	
	1 M $\Omega$ to 10 M $\Omega$	0.05 % of reading + 100 $\Omega$	
	10 M $\Omega$ to 100 M $\Omega$	0.05 % of reading + 1 k $\Omega$	
Equipment to Output AC Voltage At the listed Frequencies <sup>FO</sup>			HP 3458A
1 Hz to 40 Hz	1 mV to 10 mV	0.03 % of reading + 0.03 mV	
40 Hz to 1 kHz	1 mV to 10 mV	0.02 % of reading + 0.011 mV	
1 kHz to 20 kHz	1 mV to 10 mV	0.03 % of reading + 0.011 mV	
20 kHz to 50 kHz	1 mV to 10 mV	0.1 % of reading + 0.011 mV	
50 kHz to 100 kHz	1 mV to 10 mV	0.5 % of reading + 0.011 mV	
100 kHz to 300 kHz	1 mV to 10 mV	4 % of reading + 0.02 mV	





# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Voltage At the listed Frequencies <sup>FO</sup>			HP 3458A
1 Hz to 40 Hz	100 mV to 10 V	0.007 % of reading + 0.004 V	
40 Hz to 1 kHz	100 mV to 10 V	0.007 % of reading + 0.002 V	
1 kHz to 20 kHz	100 mV to 10 V	0.014 % of reading + 0.002 V	
20 kHz to 50 kHz	100 mV to 10 V	0.03 % of reading + 0.002 V	
50 kHz to 100 kHz	100 mV to 10 V	0.08 % of reading + 0.002 V	
100 kHz to 300 kHz	100 mV to 10 V	0.3 % of reading + 0.01 V	
300 kHz to 1 MHz	100 mV to 10 V	1 % of reading + 0.01 V	
1 MHz to 2 MHz	100 mV to 10 V	1.5 % of reading + 0.01 V	
Equipment to Output AC Voltage At the listed Frequencies <sup>FO</sup>			
1 Hz to 40 Hz	10 V to 100 V	0.02 % of reading + 0.04 V	
40 Hz to 1 kHz	10 V to 100 V	0.02 % of reading + 0.02 V	
1 KHz to 20 kHz	10 V to 100 V	0.02 % of reading + 0.02 V	
20 kHz to 50 kHz	10 V to 100 V	0.035 % of reading + 0.02 V	
50 kHz to 100 kHz	10 V to 100 V	0.12 % of reading + 0.02 V	
100 kHz to 300 kHz	10 V to 100 V	0.4 % of reading + 0.1 V	
Equipment to Output AC Voltage At the listed Frequencies <sup>FO</sup>			
1 Hz to 40 Hz	100 V to 1 000 V	0.04 % of reading + 0.4 V	
40 Hz to 1 kHz	100 V to 1 000 V	0.04 % of reading + 0.2 V	
1 kHz to 20 kHz	100 V to 1 000 V	0.06 % of reading + 0.2 V	
20 kHz to 50 kHz	100 V to 1 000 V	0.12 % of reading + 0.2 V	
Equipment to Output DC Voltage <sup>FO</sup>			
1 mV to 100 mV		0.000 9 % of reading + 0.3 $\mu$ V	
100 mV to 1 V		0.000 8 % of reading + 0.3 $\mu$ V	
1 V to 10 V		0.000 8 % of reading + 0.5 $\mu$ V	
10 V to 100 V		0.001 % of reading + 30 $\mu$ V	
100 V to 1 000 V		0.001 % of reading + 0.1 mV	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication and control Equipment used with Thermocouple Type J <sup>FO</sup>	-210 °C to -100 °C	0.26 °C	Fluke 5500A Electrical Simulation of Thermocouple Output
	-100 °C to -30 °C	0.42 °C	
	-30 °C to 150 °C	0.21 °C	
	150 °C to 760 °C	0.29 °C	
	760 °C to 1 200 °C	0.73 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type K <sup>FO</sup>	-200 °C to -100 °C	0.34 °C	
	-100 °C to -25 °C	0.18 °C	
	-25 °C to 120 °C	0.19 °C	
	120 °C to 1 000 °C	0.27 °C	
	1 000 °C to 1 372 °C	0.48 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type B <sup>FO</sup>	600 °C to 800 °C	0.44 °C	Electrical Simulation of Thermocouple Output Fluke 5500A
	800 °C to 1 000 °C	0.34 °C	
	1 000 °C to 1 550 °C	0.30 °C	
	1 550 °C to 1 820 °C	0.33 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type C <sup>FO</sup>	0 °C to 150°C	0.30 °C	
	150 °C to 650 °C	0.26 °C	
	650 °C to 1 000 °C	0.31 °C	
	1 000 °C to 1 800 °C	0.50 °C	
	1 800 °C to 2 316 °C	0.84 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type E <sup>FO</sup>	-250 °C to -100 °C	0.50 °C	
	-100 °C to -25 °C	0.16 °C	
	-25 °C to 350 °C	0.14 °C	
	350 °C to 650 °C	0.14 °C	
	650 °C to 1 000 °C	0.16 °C	





# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration Indication and Control Equipment used with Thermocouple Type N <sup>FO</sup>	-200 °C to -100 °C	0.40 °C	Electrical Simulation of Thermocouple Output Fluke 5500A
	-100 °C to -25 °C	0.22 °C	
	-25 °C to 120 °C	0.19 °C	
	120 °C to 1 000 °C	0.18 °C	
	1 000 °C to 1 372 °C	0.27 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type R <sup>FO</sup>	0 °C to 250 °C	0.57 °C	
	250 °C to 400 °C	0.35 °C	
	400 °C to 1 000 °C	0.33 °C	
	1 000 °C to 1 767 °C	0.33 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type S <sup>FO</sup>	0 °C to 250 °C	0.47 °C	
	250 °C to 1 000 °C	0.36 °C	
	1 000 °C to 1 400 °C	0.36 °C	
	1 400 °C to 1 767 °C	0.37 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type T <sup>FO</sup>	-250 °C to -150 °C	0.63 °C	
	-150 °C to 0 °C	0.24 °C	
	0 °C to 120 °C	0.16 °C	
	120 °C to 400 °C	0.14 °C	
Temperature Calibration Indication and Control Equipment used with Thermocouple Type U <sup>FO</sup>	-200 °C to 0 °C	0.56 °C	
	0 °C to 600 °C	0.27 °C	
Temperature Calibration Indication and Control Equipment used with RTD Pt 395, 100 $\Omega$ <sup>FO</sup>	-196 °C to 0 °C	0.05 °C	Electrical Simulation of RTD Output Fluke 5500A
	0 °C to 100 °C	0.05 °C	
	100 °C to 300 °C	0.07 °C	
	300 °C to 400 °C	0.09 °C	
	400 °C to 630 °C	0.10 °C	
	630 °C to 800 °C	0.12 °C	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration Indication and Control Equipment used with RTD Pt 3926, 100 $\Omega$ <sup>FO</sup>	-196 °C to 0 °C	0.05 °C	Electrical Simulation of RTD Output Fluke 5500A
	0 °C to 100 °C	0.05 °C	
	100 °C to 300 °C	0.07 °C	
	300 °C to 400 °C	0.09 °C	
	400 °C to 630 °C	0.10 °C	
Temperature Calibration Indication and Control Equipment used with RTD Pt 3916, 100 $\Omega$ <sup>FO</sup>	-196 °C to -190 °C	0.25 °C	
	-190 °C to 80 °C	0.04 °C	
	-80 °C to 0 °C	0.05 °C	
	0 °C to 100 °C	0.6 °C	
	100 °C to 260 °C	0.06 °C	
	260 °C to 300 °C	0.07 °C	
	300 °C to 400 °C	0.08 °C	
	400 °C to 600 °C	0.09 °C	
Temperature Calibration Indication and Control Equipment used with RTD Pt 385, 200 $\Omega$ <sup>FO</sup>	-196 °C to 100 °C	0.04 °C	
	100 °C to 260 °C	0.04 °C	
	260 °C to 300 °C	0.05 °C	
	300 °C to 400 °C	0.12 °C	
	400 °C to 600 °C	0.13 °C	
	600 °C to 630 °C	0.14 °C	
Temperature Calibration Indication and Control Equipment used with RTD Pt 385, 500 $\Omega$ <sup>FO</sup>	-196 °C to -80 °C	0.04 °C	
	-80 °C to 100 °C	0.04 °C	
	100 °C to 260 °C	0.05 °C	
	260 °C to 300 °C	0.06 °C	
	300 °C to 400 °C	0.08 °C	
	400 °C to 600 °C	0.08 °C	
600 °C to 630 °C	0.09 °C		



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration Indication and Control Equipment used with RTD Pt 385, 1 000 $\Omega$ <sup>FO</sup>	-160 °C to 0 °C	0.03 °C	Electrical Simulation of RTD Output Fluke 5500A
	0 °C to 100 °C	0.03 °C	
	100 °C to 260 °C	0.04 °C	
	260 °C to 300 °C	0.05 °C	
	300 °C to 400 °C	0.06 °C	
	400 °C to 600 °C	0.07 °C	
	600 °C to 630 °C	0.23 °C	
Temperature Calibration Indication and Control Equipment used with RTD Ni 385 120 $\Omega$ <sup>FO</sup>	-80 °C to 100 °C	0.08 °C	HP 8901B
	100 °C to 260 °C	0.14 °C	
Temperature Calibration Indication and Control Equipment used with RTD Cu 10 $\Omega$ <sup>FO</sup>	-100 °C to 260 °C	0.30 °C	HP 8901B
RF Microwave Equipment to Output Amplitude Modulated Signals At the listed frequencies <sup>FO</sup>			
50 Hz to 10 kHz Depths: 5 % to 99 %	150 KHz to 10 MHz	2.6 % of reading + 1 digit	
20Hz to 10 kHz Depths: 5% to 99%	150 KHz to 10 MHz	3.4 % of reading + 1 digit	
50Hz to 50 kHz Depths: 5% to 99%	150 kHz to 1.3 GHz	2.5 % of reading + 1 digit	
Rate: 20 Hz to 100 kHz Depths: 5% to 99%	150 kHz to 1.3 GHz	3.3 % of reading + 1 digit	
Equipment to Output Frequency Modulated Signals At the listed frequencies <sup>FO</sup>			
20 Hz to 10 kHz Dev: $\leq$ 40 kHz peak	0.25 MHz to 10 MHz	2.7 % of reading + 1 digit	
50 Hz to 100 kHz Dev: $\leq$ 400 kHz peak	10 MHz to 1.3 GHz	1.7 % of reading + 1 digit	
20 Hz to 200 kHz Dev: $\leq$ 400 kHz peak	10 MHz to 1.3 GHz	5.2 % of reading + 1 digit	



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	
Equipment to Measure Frequency <sup>FO</sup>	150 kHz to 10 MHz	7.2 Hz	HP 8901B	
	10 MHz to 100 MHz	8.3 Hz		
	100 MHz to 225 MHz	12 Hz		
	225 MHz to 1 300 MHz	27 Hz		
		10 Hz to 100 KHz	5.1 Hz	HP 5384A
		100 kHz to 50 000 KHz	11 Hz	
		50 MHz to 100 MHz	27 Hz	
		100 MHz to 225 MHz	47 Hz	
Equipment to Measure Signal Measuring At the Listed frequencies <sup>FO</sup>			6062A	
Up to 20 MHz	0.80 W to -1.87 W	0.04 W + 1.05 Hz		
100 KHz to 2 000 MHz	0.43 W to -4.23 W	0.04 W + 1.05 Hz		
Equipment to Output Distortion At the listed frequencies <sup>FO</sup>			HP 8903A Rate: 20 Hz to 100 kHz	
	50 Hz to 100 Hz	1mW to 1.259 mW		0.003 % of reading
	20 Hz to 20 KHz	-99.99 dB to 0 dB		1.2 dB
	20 KHz to 100 KHz	-99.99 dB to 0 dB		2.5 dB
Audio RMS Measure <sup>FO</sup>	0 Vrms to 7 Vrms	$5.8 \times 10^{-5}$ Vrms	Agilent 34401A	
Equipment to Measure RF Power 10 MHz to 18 GHz <sup>FO</sup>	-2.33 W to -0.67 W	0.06 W	HP 437B, HP 8481B, HP 8484A	
	-0.67 W to 0.67 W	0.04 W		

### Acoustic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Acoustical Measuring Equipment Fixed Points <sup>F</sup>	94 dB @ 1 kHz	0.5 dB	Acoustical Calibrator Quest QC-10



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calipers <sup>FO</sup>	0.1 in to 6 in (2.54 mm to 152.4 mm)	(478.67 + 1.34L) $\mu$ in [(12.16 + 1.34 x 10 <sup>-3</sup> L) $\mu$ m]	Gage Blocks
Digital and Dial Indicators <sup>FO</sup>	0.1 in to 2 in (2.54 mm to 50.8 mm)	(59.84 + 74.01 x 10 <sup>-2</sup> L) $\mu$ in [(1.52 + 1.88 x 10 <sup>-2</sup> L) $\mu$ m]	
Depth Gages <sup>FO</sup>	0.1 in to 6 in (2.54 mm to 152.4 mm)	(478.67 + 1.34L) $\mu$ in [(12.2 + 1.34 x 10 <sup>-3</sup> L) $\mu$ m]	
Thickness Gages <sup>FO</sup>	0.1 in to 0.5 in (2.54 mm to 12.7 mm)	(59.84 + 2.91 x 10 <sup>-2</sup> L) $\mu$ in [(1.52 + 7.40 x 10 <sup>-4</sup> L) $\mu$ m]	
Gage Blocks <sup>F</sup>	0.1 in to 20 in (2.54 mm to 508 mm)	(3 + 1.7L) $\mu$ in [(0.076 + 0.001 7L) $\mu$ m]	Gage Blocks Comparator Master Gage Blocks b Grade 00
Optical Comparator X & Y Axis Linearity <sup>O</sup>	Maximum 10 in Maximum 10 in	(370 + 60L) $\mu$ in	Glass Scales
Optical Comparator Angularity <sup>O</sup>	0° to 180°	0.2°	Glass Scales
Squareness <sup>F</sup>	4 in of Y axis travel or maximum Y axis travel if maximum is less than 4 in.	300 $\mu$ in	Master square
Surface Plates Flatness <sup>O</sup>	16 in to 50 in	98 $\mu$ in	Planekator
CMM Repeatability <sup>O</sup>	1 in	94 $\mu$ in	-Sphere 1 in

### Times & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure Frequency <sup>FO</sup>	0.01 Hz to 11.999 kHz	0.086 % of reading	Fluke 5500A
	12 kHz to 2 MHz	0.086 % of reading	
Equipment to Output Frequency <sup>FO</sup>	1 Hz to 40 Hz	0.05 % of reading	HP 3458A
	40 Hz to 10 MHz	0.01 % of reading	
Time Interval <sup>FO</sup>	0.35 ms to 5 s	0.10 ms/s	Tektronix TDS 3032

### Mass, Force & Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Analytical Balance <sup>O</sup>	1 g to 100 g	(9 x 10 <sup>-4</sup> + 2 x 10 <sup>-4</sup> Wt) g	Weight Set 1 g to 100 g, 500 g to 3 Kg, 5 Lbs to 250 Lbs NIST Handbook 44 Class F
Electronic Balance - Counting Balance <sup>O</sup>	500 g to 3 kg	(2.10 x 10 <sup>-4</sup> + 6.866 x 10 <sup>-8</sup> Wt) g	
	5 lb to 250 lb	(3 x 10 <sup>-4</sup> + 1 x 10 <sup>-4</sup> Wt) lb	





# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva #07, Fracc. Conquistadores, Cd. Rio Bravo, Tamps. 88993

Contact Name: Luis Rivera Phone: 899-263-9910

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gauges <sup>FO</sup>	0.1 MPa to 2.413 1 MPa (14.5 psi to 349.99 psi)	5.6 kPa (0.809 psi)	Fluke PV 350
Vaccum Gauges <sup>FO</sup>	-99.898 kPa to 0.102 kPa (-29.5 inHg to 0.030 inHg)	7.6 kPa (2.244 inHg)	
Pressure Gauges <sup>F</sup>	100 psi to 10 000 psi	1.4 % of reading	Dead Weight Tester Ametek R-55
Indirect Verification of Rockwell Hardness Testers HRC <sup>O</sup>	20 HRC to 39 HRC	0.37 HRC	ASTM E18-08A and Calibrated Rockwell hardness Test Blocks
	40 HRC to 59 HRC	0.35 HRC	
	60 HRC to 70 HRC	0.32 HRC	
Torque Wrench <sup>FO</sup>	0.1 N•m to 10 N•m	0.62 % reading	Torque Analyzer Crane W/ Load Cell Rotary
Electronic Screwdriver <sup>FO</sup>	0.1 N•m to 10 N•m	0.62 % reading	
Torque Tester Analog / Digital <sup>F</sup>	20 lbf•in to 400 lbf•in	0.44 lbf•in	Weights Class F and 4 in Torque Wheel

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this calibration onsite at the customer's location.
6. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
7. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.